INFRARED BODY THERMOMETER

1 pcs/box
40 pcs/carton

3 color backlight as fever reminder

Green：Normal

≤37.3℃(99.1℉)

Orange：Low fever

37.4℃ (99.32℉)-
38℃ (100.4℉)

Red：High fever

≥38℃ (100.4℉)

Advantages

- High accuracy
- Fast response
- One-click design
- Ultra-light weight

One-click design

Easy-use design: One-click, Two Model
For all age: Adult,Child,Infant

Product description

This product is mainly composed of infrared probe assembly, main circuit board assembly, LED display assembly and shell assembly

1.Infrared probe

2.LED screen

3.Setting Key

4.Measurement key

5.Battery cover

6.Holding case

Install the battery
Press the arrow to open the battary cover

3 Steps for Temperature

1. Place the Infrared Thermometer in a tested environment for 30 minutes
2. Make sure the forehead is not covered by hair, perspiration, cap or scarf
3. Keep the measuring distance as 3 cm

Features

Fast measuring reaction:≤2 sec
Measuring Distance: 2cm~5cm
Memory Data: 100 sets
Auto power-off: ≤15sec
Transport storage Temperature: -20°C ~ 55°C
Relative humidity: ≤95%
Power supply: DC 3V (2x "AAA" batteries)
Body Temperature Range: 34°C~42.9°C
Measurement accuracy: 34°C~34.9°C : ±0.3°C
35°C~42°C : ±0.2°C
42.1°C~42.9°C : ±0.3°C
LED display with 3 colors Backlight:Green: 34°C~37.3°C
Red: 38.1°C(default)~42.9°C
Storage: Store in a dry , cool, sunless place

Certificate

CERTIFICATION OF REGISTRATION

This certifies that:
Longnan Renzhong Medical Equipment Co., Ltd
Information Industry Technology City, Longnan Economic Development Zone,
Longnan, Ganzhou, Jiangxi province, CN 341700

is registered and has listed the following medical device with the U.S. Food and Drug Administration:

Owner/Operator Number : 10063771
Listing Number: D077111
Product Code : KFO
Product : Thermometer
Model(s) : YK001
Date Of Registration Status: 2020

Issued: Mar. 23, 2020
Validity Period 2020-12-31

Shenzhen POCE Technology Co.,Ltd.
H Building, Hongfa Science and Technology Park, Tangtou, Shiyan, Bao'an District, Shenzhen, China

CERTIFICATE OF CONFORMITY

Certificate No.	:	POCE20011705ICCR
Applicant	:	Longnan Renzhong Medical Equipment Co., Ltd
Address	:	Information Industry Technology City, Longnan Economic Development Zone, Longnan, Ganzhou, Jiangxi province, China
Manufacturer	:	Longnan Renzhong Medical Equipment Co., Ltd
Address	:	Information Industry Technology City, Longnan Economic Development Zone, Longnan, Ganzhou, Jiangxi province, China
Product	:	INFRARED BODY THERMOMETER
Trade Name	:	AFK
Model(s)	:	YK001, YK002, YK003, YK004
Test Report No.	:	POCE20011705ICRR
Test Standards	:	IEC 62321-3-1:2013; IEC 62321-4:2013; IEC 62321-5:2013; IEC 62321-6:2015; IEC 62321-7-1:2015; IEC 62321-8:2017

The EUT described above has been tested by us with the listed standards and found in compliance with the council **RoHS Directive(EU) 2015/863 amending Annex II to Directive 2011/65/EC**. It is possible to use **CE** marking to demonstrate the compliance with this **RoHS** Directive.

Date: Mar. 19, 2020

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production, and other relevant directives have to be observed.

Web: http://www.poce-cert.com Tel: +86-755-29113202 E-mail: service@poce-cert.com

Shenzhen CTO Technology Service Co.,Ltd
9 I F, Block B, 196 Tangtou Avenue, Langtou Community, Shiyan Street, Baoan District, Shenzhen China.

CERTIFICATE OF CONFORMITY

Certificate No.	:	CTO20031807401CS
Applicant	:	Longnan Renzhong Medical Equipment Co., Ltd
Address	:	Information Industry Technology City, Longnan Economic Development Zone, Longnan, Ganzhou, Jiangxi province, China
Manufacturer	:	Longnan Renzhong Medical Equipment Co., Ltd
Address	:	Information Industry Technology City, Longnan Economic Development Zone, Longnan, Ganzhou, Jiangxi province, China
Product	:	INFRARED BODY THERMOMETER
Trade Name	:	AFK
Model(s)	:	YK001, YK002, YK003, YK004
Test Report No.	:	CTO20031204RDRS CTO20031205ERS EN 60601-1: 2006 + A1: 2013+A12:2014
Test Standards	:	EN 60601-1-11:2015 EN 60601-1-2:2015

The EUT described above has been tested by us with the listed standards and found in compliance with the council **Medical Devices Directive(MDD) 93/42/EEC**. It is possible to use **CE** marking to demonstrate the compliance with this **MD Directive**

Date: Mar. 19, 2020

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production and other relevant directives have to be observed.

Web: www.cto-cert.com E-mail: service@cto-cert.com